



EUROPEAN UNION

Delegation of the European Union to Bosnia and Herzegovina

Sarajevo,

Corrigendum no.1

Development of Infrastructure of Metrology system in Bosnia and Herzegovina

Publication reference: EuropeAid/130697/C/SUP/BA
Tender No: EC/BiH/TEN/10/040

Following changes in **Annex II, Technical Specifications** and **Annex III, Technical offer** have been made, as listed below:

LOT 2: Time and frequency standards and instruments

Item 2.4 Phase Comparator

The following text :

Supply voltage AC: (90 – 265) V AC,

Is REPLACED :

Supply voltage AC: (90 – 264) V AC,

Item 2.11 High Performance Timer/Counter

The following text :

Rise time: about 1ns.

Is REPLACED :

Rise time: 1ns \pm 10%.

All other terms and conditions of the Tender Dossier remain unchanged.