



EUROPEAN UNION

DELEGATION TO BOSNIA AND HERZEGOVINA

Supply of equipment for phytosanitary laboratories
EuropeAid/133138/C/SUP/BA; Tender no: EC/BIH/TEN/12/029

CORRIGENDUM No. 1

- 1. Following change has been made in Annex II + III, Technical Specifications, Lot 1- Equipment for Phytosanitary Laboratories**

Item 1.6 Glassware Drying Cabinet

Requirement

- *Capacity of the cabinet minimum 500/700 litres*

Shall be replaced with:

Internal capacity of the cabinet: minimum 500 L to maximum 750 L

Item 1.14 Bench top (table) Centrifuge:

Requirement

- *Included rotors (fixed angle and swing-out) with adaptors for 15 mL, 50 mL and 100 mL falcon tubes (tubes included)*

shall be replaced with:

Included rotors (fixed angle and swing-out) with adaptors for 15 mL, 50 mL (falcon tubes), and 100 mL tubes (tubes included).

Quantity: 50 Tubes for each dimension

Item 1.16 Real – Time PCR cycler:

Requirement

- *Heating rate: 4 °C or higher*

shall be replaced with:

Heating rate: 2 °C or higher

Item 1.20 Lyophilizer

Requirements

- *8 port manifold with valve (1pcs.) and one distributor per port for ampoules*
- *At least 10 port manifold (1pcs.) and one distributor per port for ampoules*

shall be replaced with:

8 port manifold with valve (1pcs.) and one distributor for ampoules (or equivalent system).

4. All other terms and conditions of the Tender Dossier remain unchanged.

Emile Mabita

Head of Contracts, Finance and Audit Section