

EUROPEAN UNION

DELEGATION TO BOSNIA AND HERZEGOVINA

Supply of equipment for phytosanitary laboratories EuropeAid/133138/C/SUP/BA; Tender no: EC/BIH/TEN/12/029

CORRIGENDUM No. 2

1. Following changes have been made in Annex II + III, Technical Specifications, Lot 1, Equipment for Phytosanitary Laboratories:

Item 1.1 Sample processing equipment

Requirement: Speed from at least 300 to 3000 rpm

Shall be replaced with: Speed range: minimum 500 rpm, maximum above 3000 rpm

Item 1.22.2 Seed sampling stick

Requirements: Diameter of the hole around 32 mm Length: not more than 600/700 mm

Shall be replaced with: Diameter of the hole 32 mm +/- 3 mm Length: not more than 1000 mm

Item 1.3 Bench top centrifuge

Requirement: *Standard power supply: 220V-50 Hz*

Shall be replaced with: Standard power supply: 230 V - 50/60 Hz

Item 1.7 Transport containers for waste Requirement:

With accessories

Shall be replaced with: Supplied with a minimum 100 bags that fit in the respective container

Item 1.9 Laboratory freezer

Requirements: Working temperature: -30°C Power supply: 220/230V, 50 Hz Shall be replaced with: Working temperature: -28°C Standard power supply: 230 V - 50/60 Hz

Item 1.19 Ultra Deep Freezer

Requirement: Volume (capacity): at least 450 litres

Shall be replaced with: Volume (capacity): at least 445 litres

Item 1.20 Lyophilizer

Requirement: Power supply 220 to 230 220 V, 50-60Hz

Shall be replaced with: Power supply: 230 V - 50/60 Hz

Item 1.21 Vertical Laminar air flow

Requirement: EN 12469-2000 certified

Shall be replaced with: Complies with EN 12469-2000

Item 1.31 Laboratory Refrigerator

Requirements: Volume from 300 up to 350 L Minimum temperature range +4 to +15°C

Shall be replaced with: Volume from 300 up to 355 L Minimum temperature range +4 to +10°C

Item 1.34 Water Distiller Requirement: Power supply: 220 V 50/60 Hz

Shall be replaced with: Power supply: 230 V - 50/60 Hz

2. All other terms and conditions of the Tender Dossier remain unchanged.

Emile Mabita Head of Contracts, Finance and Audit Section