



CORRIGENDUM No. 1 to the Tender Dossier
Improving capacity of Indirect Taxation Authority
EuropeAid/139007/DH/SUP/BA; Tender No: EC/BiH/TEN/17/015

The Tender Dossier is corrected as follows:

ANNEX II + III: TECHNICAL SPECIFICATIONS + TECHNICAL OFFER

1) LOT N° 3: Optical laboratory instruments, instruments for alcohol determination, and for fuel testing

Item 3.1. Cloud and Pour Point Analyzer with built-in cooling

(a) The requirement:

Way of detection:

Pour point detection optically, tilting method

now reads:

Pour point detection optically, tilting method or micro thermal detection, tilting method.

2) Item 3.2. FTIR (NIR/MID IR) spectrometer for octane and cetane number determination & characterization of gasoline and diesel fuels

(a) In the requirement: "Scanning in the range of- 450 to 7000 cm-1, as minimum

now reads:

"Scanning in the range of- 600 to 4000 cm-1, as minimum"

(b) In the requirement: " other oxygenates (0-25 v %);

now reads:

other oxygenates (0-20 v%);

c) In the requirement: "Sample volume up to 20 ml"

" now reads:

"Sample volume up to 25 ml"

d) In the requirement: "Auto sampler" now reads:

"Automatic sampler on/or automatic injector"